

# PTS 2.3 genX

**Three-phase, fully automatic test system with a class 0.1 reference standard and integrated three-phase current source**



The PTS 2.3 genX portable test system consists of an integrated three-phase current source unit and a three-phase electronic reference standard of accuracy class 0.1. Characteristic features of the PTS 2.3 genX are its wide measuring range, high accuracy and high tolerance to unwanted external influences.

The PTS 2.3 genX allows the monitoring of meter installations as well as analysis of the local mains conditions.

#### Advantages

- Easy verification of meters under precise load conditions, using the built-in, compact, current source
- Automatic operation with predefined load points without the need for an external PC
- Large 9" (800 x 480 pixels) TFT touch screen colour display with graphical user interface
- Data transfer and communication via USB (Type B), ETHER-NET or WLAN
- Built in web server for remote display of graphical user interface and remote control of the unit
- Data storage on removable SD memory card
- Two USB (type A) connectors for connection of peripherals like mouse, keyboard

#### Functions

- Independent generation of single- or three-phase current loading conditions for verification of meters using the incoming supply voltage
- Active, reactive and apparent energy measurement for three phase, 3 or 4-wire, systems with integrated error calculator and pulse output
- Vector diagram, harmonics spectrum, wave form and rotary field display for analysis of the mains conditions
- Burden measurement of Current Transformer (CT) and Potential Transformer (PT)
- Ratio testing of Current Transformers (CT)

#### Options

- Software CAlegation
- UCT 120.3 set of 3 clamp-on CT's 120 A
- UCT 1000.3 set of 3 clamp-on CT's 1000 A
- UCT LEM.3 set of 3 flexible current probes FLEX 3000 (30/300/3000A)
- UCT AMP-LiteWire 3-phase adapter set for AmpLiteWire + primary high voltage current sensor AmpLiteWire 2000 A
- UCT VOLT-LiteWire 3-phase adapter set for VoltLiteWire + primary voltage sensor VoltLiteWire 40 kV

# Technical Data PTS 2.3 genX

## General

Auxiliary power supply:	Power may be taken from the auxiliary supply or the measuring circuit at: 88 VACmin ... 264 VACmax / 47 ... 63 Hz Protection: Switch off supply at > 276 VAC
Power consumption:	max. 400 VA
Housing:	Hard Plastic
Dimensions:	W 450 x D 297 x H 188 mm
Weight:	approx. 14.5 kg
Operation temperature:	-10 °C ... +50 °C
Storage temperature:	-20 °C ... +60 °C
Relative humidity:	≤ 85% at Ta ≤ 21°C ≤ 95% at Ta ≤ 25°C, 30 days / year spread

## Safety

Isolation protection:	IEC 61010-1:2010
Measurement Category:	300V CAT III, 600V CAT II
Degree of protection:	IP-68 (housing closed) IP-40 (housing open)

## Current Source

Current Range	1 mA ... 120 A	
Output power	60 VA (per phase)	
	Internal Ranges	S <sub>max</sub> / U <sub>max</sub>
	1 mA ... 12 mA	60 mVA / 5 V
	12 mA ... 120 mA	600 mVA / 5 V
	120 mA ... 1.2 A	6 VA / 5 V
	1.2 A ... 12 A	60 VA / 5 V
	12 A ... 80 A	60 VA / 0.75 V
	80 A ... 120 A	60 VA / 0.5 V
Resolution	0.01 % of end of internal range	
Accuracy	≤ 0.1 % of end of internal range	
Distortion Factor	≤ 0.25 % on linear load	
Stability	≤ 0.03 % (30 min.) ≤ 0.1 % (1 h)	
Load Regulation	≤ 0.01 % (from 0 % ... 100 % load)	
Power Factor of Load	1 – 0.1 ind.	
Bandwidth	30 Hz ... 3 kHz (-3 dB)	
Phase Angle	Range	Accuracy
	-180° ..+180°	± 0.1°
		0.01°
Frequency	Range	Accuracy
Mode Sync (to input voltage)	40 Hz-70 Hz	
Mode Frequency	40 Hz-70 Hz	± 0.01 Hz
Efficiency	> 85 %	

## Reference Standard - Measurement Range

Measuring Quantity	Range	Input / Sensor
Voltage (phase - neutral)	5 V ... 500 V 10 mV ... 5 V	U1, U2, U3 U1 (Burden)
Current	1 mA ... 12 A 1 mA ... 120 A 10 mA ... 120 A 100 mA ... 1000 A 3 A ... 3000 A	12 A (I1, I2, I3) 120 A (I1, I2, I3) UCT 120.3 UCT 1000.3 FLEX 3000

## Reference Standard - Measurement Accuracy

Voltage / Current	≤ ± E [%] <sup>1,2,4</sup>
Measuring Quantity	Cl. 0.1
Voltage (U1, U2, U3, N)	0.1 0.2
Current direct 12 A / 120 A	0.1 0.1
Current CT 120A UCT 120.3	0.2 1.0
Curr. CT 1000A UCT 1000.3	0.2 1.0
Current FLEX 3000 UCT LEM.3	0.1 + E <sub>M</sub>
Burden Voltage (U1)	0.5 0.5
	10 mV ... 500 mV

Power / Energy	Voltage: 30 V ... 500 V (U - N)	≤ ± E [%] <sup>1,2,3</sup>
Measuring quantity / Input I	Range	Cl. 0.1
<b>Active (P), Apparent (S) and Reactive (Q) Power / Energy</b>		
Direct 12 A / 120 A (I1, I2, I3)	10 mA ... 120 A 1 mA ... 10 mA	0.1 0.1
Current CT 120A UCT 120.3	100 mA ... 120 A 10 mA ... 100 mA	0.2 1.0
Curr. CT 1000A UCT 1000.3	10 A ... 1000 A 1 A ... 10 A	0.2 1.0

Influence of external magnetic fields (45 Hz ... 66 Hz):  $\leq 0.07 \% / 0.5 \text{ mT}^3$

Temperature coefficient (TC):	Range	≤ ± TC [%/°C] <sup>3</sup>
	0° C ... +40°C	0.005
	-10° C ... +50°C	0.008

Frequency / Phase Angle / Power Factor	≤ ± E
Measuring Quantity	Range
Frequency (f)	40 Hz ... 70 Hz
Phase Angle (φ)	0.00 ° ... 359.99 °
Power Factor (PF)	-1.000 ... +1.000

CT/PT Ratio	≤ ± E [%] <sup>1,2</sup>
Ratio error E: Sum of errors of inputs used for primary (IP, UP) and secondary (IS, US) measurements.	E <sub>P</sub> + E <sub>S</sub>

CT/PT Burden	≤ ± E[%] <sup>1,2</sup>
Operating burden S <sub>n</sub> : Sum of errors of inputs used for voltage (U) and current (I) measurement.	E <sub>U</sub> + E <sub>I</sub>

## Notes

- 1 x.x :Related to the measuring value
- 2 x.x :Related to the measuring range final value (full scale, FS), E(M) = FS/M \* x.x (e.g. 0.1 at FS = 120 mA, E(20mA) = 120/20 \* 0.1 = 0.5 %)
- 3 Fundamental frequency in the range 45 ... 66 Hz
- 4 S: x.x, P,Q: x.x / PF (related to apparent power), 3- and 4-wire networks
- 5 Typical values, determined on the basis of monthly calibrations and calculated by least square method

**Pulse Input / output** Input 1 can be configured as output

Input level:	4 ... 12 VDC (24 VDC)
Input frequency:	max. 200 kHz
Supply:	12 VDC (I < 60 mA)
Output level:	5V
Pulse length:	≥ 10 µs
<b>Meter constant:</b>	C = C <sub>0</sub> / (In * Un) C <sub>0</sub> = 60'000'000 [imp/Wh(varh,VAh)] The meter constant depends on the highest selected internal ranges In, Un. The actual constant CPZ <sub>1</sub> with unit [imp/Ws (vars, VAs)] is indicated on the display at frequency output.
Internal current ranges In [A]	
Direct 12/120 A (I1, I2, I3)	0.004    0.012    0.04    0.12
	0.4    1.2    4    12
	40    80    120
Current CT 120A UCT 120.3	0.12    1.2    12    120
Current CT 1000A UCT 1000.3	1    10    100    1000
FLEX 3000 UCT LEM.3	-    30    300    3000
Internal voltage ranges Un [V]	
Burden U1	0.5    5
Direct U1, U2, U3, N	250    500
Example: Un = 250 V, In = 12 A C = 20'000 [imp/Wh(varh,VAh)]	
Output frequency:	CPZ <sub>1</sub> = C / 3'600 [imp/Ws(vars, VAs)] f <sub>0</sub> = CPZ <sub>1</sub> * Σ(Q <sub>S</sub> , S <sub>S</sub> ) f <sub>max</sub> = CPZ <sub>1</sub> * 3 * Un * In = 5.55555 imp/Ws * 3 * 250V * 12A = 50'000 [imp/s] Factor 3 for 3-phase system